

Evonik Degussa GmbH 45764 Marl, Germany

July 18, 2011

To our customers

Dr. Hartwig Lange
Coatings & Additives
Environment, Health, Safety &
Quality
Paul-Baumann-Straße 1
45764 Marl
Phone +49 2365 - 494459
Fax +49 2365 - 497275
Hartwig.Lange@evonik.com

**VESTAMID® Care ML21
USP VI**

Dear Sir or Madam,

We have performed studies with

- VESTAMID® Care ML21

to demonstrate that this VESTAMID® Care ML grade meets the requirements of USP Class VI (70°C). With VESTAMID® Care ML21 the following GLP compliant studies have been done:

- Systemic injection test - i.p. administration
- Systemic injection test - i.v. administration
- Intracutaneous test
- Implantation test

We are happy to inform you that based on these studies VESTAMID® Care ML21 meets the requirements of USP Class VI (70°C).

Yours faithfully

 Evonik Degussa GmbH


ppa. Dr. Jörg Limper


i. V. Dr. Rolf Ebert
EUROTOX certified Toxicologist

Evonik Degussa GmbH
Rellinghauser Straße 1-11
45128 Essen
Germany
Phone +49 201 177-02
Fax +49 201 177-3475
www.evonik.com

Supervisory Board
Dr. Klaus Engel, Chairman
Board of Management
Patrik Wohlhauser, Chairman
Dr. Thomas Haeberle, Thomas Wessel

Registered Office is Essen
Register Court
Essen Local Court
Commercial Registry B 20227